



Specification

1.01691.0025 Canada balsam for microscopy

Specification		
Appearance	nearly clear, yellow resinous	
Refractive index (n 20/D)	1.515 - 1.530	
Turbidity	≤ 2	NTU
Acid value	75 - 90	
Loss on drying (120 °C)	≤ 35	%
Suitability for microscopy	passes test	

Dr. Timm Schlegelmilch

Responsible laboratory manager quality control

This document has been produced electronically and is valid without a signature.