

Corning Incorporated Life Sciences Product Description



Catalog Number: 9610-100

Product Description: PYREX® Tube, Weathering Test, Graduated, 100 mL

Component Materials:

Type I Class A - Low-Expansion Borosilicate Glass, per ASTM E438 (Standard Specification for Glasses in Laboratory Apparatus)

Attributes:

Capacity - 100 mL \pm 1.0 mL

Product Dimensions - 37 mm OD x 200 mm High

Vent Hole - 3 mm Diameter

Autoclaving:

This product will withstand repeated standard autoclave cycles of 121°C at 15 psi for 15 minutes.

Heating and Cooling:

This product is primarily designed for a specific use that does not involve direct heating. In normal use, temperatures will be from - 50°C to +125°C. This is below the temperature at which thermal failure could be a problem.

Vacuum / Pressure:

This product is not designed for and is not rated for vacuum or pressure applications.

Sterilization:

Shipped Not Sterile.

TSE/BSE:

Materials of animal origin are not used in the manufacture of this product.

Lot Number Designation:

Each production lot is uniquely identified with an 8 digit lot number as follows: First 3 digits - Julian date, start of manufacturing; Next 2 digits - Year of manufacture; Last 3 digits - Batch identification.

Certifications:

ISO 9001

To get Certificates per Lot and Serialized PYREX® items, visit www.corning.com/resources.

Corning Glassware page:



Enter item ID and serial or lot number of
your PYREX product.

